

LOCATION	WRITTEN BY:	APPROVED BY:	DATE CREATED	LAST REVISION
PTSD Schools	L. Carriere /N. Dyck	N. Dyck	March 1, 2013	New

HAZARDS PRESENT	PERSONAL PROTECTIVE EQUIPMENT (PPE)	ADDITIONAL REQUIREMENTS
✓ Crouching / Bending / Kneeling	✓ Universal Precautions	<ul style="list-style-type: none"> ✓ Equipment orientation ✓ PPE Awareness


SAFE WORK PROCEDURE

NOTE:

- The Good Samaritan Protection Act (Bill 214) protects a person who voluntarily provides emergency medical services, aid or advice to a victim of an accident or medical emergency at the immediate scene of the accident or emergency. **No liability is posed unless there is clear evidence of gross negligence.**
- Adult battery/pad pack is to be used on patients eight (8) years old and older and weigh more than 25 Kg (55lbs).
- Pediatric battery/pad pack is to be used on patients younger than eight (8) years old and weigh less than 25 Kg (55lbs).

PROCEDURE:

1. Upon arriving at the scene, ensure the area is clear of hazards and is safe.
2. Check to see if the person is non-responsive by pinching the shoulders and shouting their name or "Are you OK".
3. If no response, have someone call 911 or 9-911 depending on your building's phone system
4. Watch, feel and listen for signs of life by opening the airway and check for breathing.
5. Have another person retrieve the AED or call the main office and have someone retrieve and bring the AED to the room (Note: identify your location / room #).
6. If the person is non-responsive (breathing and pulse are absent or irregular) begin chest compressions (at least 100 / minute for 2 minutes).
7. Once the AED arrives, press the on/off button and wait for instructions. You will hear voice prompts and see visual cues coming from the machine.
8. Do not touch the patient as the AED is analyzing.
9. After delivery of 1 shock, resume chest compressions for 2 minutes. The AED will automatically reanalyze.
10. Once EMS personnel arrive they will assume responsibility for the patient and retrieve the information stored on the AED if required.



REGULATORY REQUIREMENTS

- WS&H Act W210, Section 4, 5, 7, 7.1
- Defibrillator Public Access Act & Regulations
- Manufacturer's Instructions for use